

BLANK PAGE





Indian Standard

SPECIFICATION FOR GIMLET FOR TREPHINE

- 1. Scope Govers requirements for gimlet for trephine used in veterinary surgery.
- 2. Shape and Dimensions As shown in Fig. 1.




Fig. 1 Gimlet For Trephine

- 2.1 A deviation of ± 2.5 percent shall be allowed on all dimensions.
- 3. Materials Gimlet and handle shall be made of stainless steel conforming to Designation 30Cr13 or 40Cr13 of IS: 6603 1972 'Specification for stainless steel bars and flats'.
- 4. Requirements The gimlet shall have sharp, well cut threads and shall be screwed into the handle and countersunk flush with the handle surface. The surface of the gimlet shall be free from burrs, pits, cracks and other surface defects. The gimlet shall be passivated and polished bright.
- 5. Hardness The gimlet shall be uniformly hardened to a hardness of 500 to 600 HV. The hardness shall be tested in accordance with IS: 1501-1968 'Method for Vickers hardness test for steel (first revision)'.
- 6. Functional Test When moderate manual pressure is applied to the gimlet, it shall effectively drill a hole in a fresh piece of bone 25 mm thick. The test shall be carried out six times and on completion of the test, the threads of the gimlet shall show no sign of damage.
- 6.1 Corrosion Resistance Test Shall conform to IS: 7531-1975 'Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical interuments'.
- 7. Marking The gimlet shall be legibly and indelibly marked with the manufacturer's name, initials or recognized trade-mark, and the words 'stainless steel or letters 'SS'.
- 7.1 /S/ Certification Marking Details available with the Indian Standards Institution.
- 8. Packing As agreed to between the purchaser and the supplier.
- Sampling As agreed to between the purchaser and the supplier.

Adopted 23 July 1984

© October 1984, ISI

Gr 1